

Certificate of Analysis - Reagent RVD-E

<u>Product Name:</u>	Reagent RVD-E	<u>Date of Manufacture:</u>	December 2020
<u>Catalog #:</u>	7833025 / 7833100 / 7833500	<u>Lot #:</u>	0004
<u>Product Line:</u>	PCRopsis TM	<u>Expiration Date:</u>	June 2022
		<u>Storage Conditions:</u>	4 - 25°C
		<u>Handling Conditions:</u>	Keep sterile

Testing Characteristic

Test Result

Lot

Compare Ct values between Reagent RVD-E + RT-qPCR to RNA extraction + RT-qPCR from stock sample

Observed Ct relative to RNA extraction:
+2.9

Pass: +/- 5Ct:
PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: December 2020

This certificate of analysis verifies the above-mentioned lot of Reagent RVD-E conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto
Name of Approving Manager


Signature of Approving Manager